

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2057459	(X3) Date Survey Completed 01/15/2019
Name of Provider or Supplier Dermatology Specialists, The	Street Address, City, State 176 Mountain Avenue, Hackettstown, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual (PM) and interview with Laboratory Staff (LS), the laboratory failed to follow the PM for microscope maintenance from 3/7/17 to the date of the survey. The finding includes: 1. The PM stated microscope maintenance was to be performed every 2 years but documentation revealed the maintenance was last performed on 3/24/15. 2. The LS confirmed on 1/16 /19 at 1:30 pm that the laboratory did not follow the PM.</p>