

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D2070721	<b>(X3) Date Survey Completed</b>  01/18/2018
<b>Name of Provider or Supplier</b>  Coyle S Connolly, Do Pa	<b>Street Address, City, State</b>  2099 New Albany Road, Cinnaminson, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: a) Based on surveyor review of the Procedure Manual (PM) and interview with the Office Manager (OM), the laboratory failed to have a policy for the retention of histopathology slides from 12/14/15 to the date of survey. The PM confirmed on 1/18 /18 at 10:00 am that the OM did not have the policy. b) Based on surveyor review of the Biannual Assessment (BA) procedure, PM and interview with the OM, the laboratory failed follow the BA The finding includes: 1.The BA procedure stated that the BA will be performed every six months. 2. BA was performed once in the calendar year 2017 3. The OM confirmed on 1/18/18 at 10:10 am that the BA procedure was not followed.</p>