

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2100417	(X3) Date Survey Completed 03/26/2024
Name of Provider or Supplier Dermatology Clinic, The	Street Address, City, State 901 West Main Street, Suite 201, Freehold, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of Histology reagents and interview with the Office Manager (OM), the laboratory failed to discard expired Histopathology reagent from 12/31/2021 to the date of survey. The findings include: 1. Adventek Tissue marking dyes were expired as follows: a) Blue marking day lot # 083812 expired 6/1/2021 b) Black marking dye lot #092020 expired 12/31/2021 c) Red marking day lot # 083863 expired 6/1/2021 2. Approximately 200 patients were tested with expired reagent. 3. The OM confirmed on 3/26/24 at 10:30 am that the laboratory used expired reagents.</p>