

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  31D2122083	<b>(X3) Date Survey Completed</b>  02/21/2019
<b>Name of Provider or Supplier</b>  Patient First - Voorhees	<b>Street Address, City, State</b>  705 Haddonfield-Berlin Road, Voorhees, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6051</b>	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)(v)</p> <p>The procedures for evaluation of the competency of the staff must include, but are not limited to assessment of test performance through testing previously analyzed specimens, internal blind testing samples or external proficiency testing samples.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Competency Assessment (CA) records and interview with the Technical Consultant (TC), the TC failed to document CA on assessment of test performance through testing unknown samples for five of five testing personnel for Hematology, Serum Human Chorionic Gonadotropin and Wet Mount tests in 2018. The TC # 2 listed on CMS form 209 confirmed on 2/21/19 at 10:00 am that CA was not assessed on above procedure.</p>