

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 31D2131425	<b>(X3) Date Survey Completed</b> 06/06/2019
<b>Name of Provider or Supplier</b> Associates In Digestive Diseases Pa	<b>Street Address, City, State</b> 25 Morris Avenue, Springfield, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Procedure Manual, observation of microscope and interview with the Clinical Consultant (CC), the laboratory failed to follow Microscope Maintenance Procedure (MMP) in the calendar year 2018. The finding includes: 1. The PM stated to perform preventive maintenance of microscope annually but there was no evidence of maintenance performed. 3. The CC confirmed on 6/6/19 at 1:00 pm that the laboratory did not follow MMP.</p>