

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 31D2151571	<b>(X3) Date Survey Completed</b> 04/05/2019
<b>Name of Provider or Supplier</b> Nj Certified Dermatology, Pc	<b>Street Address, City, State</b> 271 South New Prospect Rd, Jackson, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6084</b>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1445(e)(2)</p> <p>The laboratory director must ensure that the physical plant and environmental conditions provide a safe environment in which employees are protected from physical, chemical, and biological hazards.</p> <p>This STANDARD is not met as evidenced by: Based on the lack of a Flammable Cabinet, observation of flammable reagents, and interview with the General Supervisor (GS) the Laboratory Director (LD) failed to provide a safe environment from physical and chemical hazards from September 2018 until the date of survey. 1) Xylene and alcohol reagents were kept in a cabinet under a sink. 2) Xylene and Alcohol reagents are "fire and explosion hazards" as stated by the manufacture. 3) The GS confirmed at 1:00 pm on 4/5/19 that the LD failed to provide a safe environment from physical and chemical hazards.</p>