

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2154902	(X3) Date Survey Completed 12/07/2021
Name of Provider or Supplier Schweiger Dermatology Group, Pllc - Morristown	Street Address, City, State 290 Madison Avenue, Building 5, Suite 9, Morristown, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the Biannual Assessment (BA) records and interview with the Office Manager (OM), the laboratory failed to verify the accuracy of Histopathology testing twice annually in the calendar year 2021. The OM confirmed on 12/9/21 at 12:15 pm that the laboratory did not perform BA for Histopathology testing twice in 2021.</p>
D6102	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(12)</p> <p>The laboratory director must ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of Personnel Records (PR) and interview with the Office Manager (OM), the Laboratory Director failed to ensure that the education records were available for all Testing Personnel (TP) on the date of the survey. The finding includes: 1. Education records were not available for one out of one TP. 2. The OM confirmed on 12/7/21 at 12:00 pm that education records were not available.</p>