

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2157832	(X3) Date Survey Completed 04/15/2025
Name of Provider or Supplier Monmouth Medical Center Vantage Point	Street Address, City, State 100 Route 36 East, Suite 1m, West Long Branch, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6020	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(5)</p> <p>(e)(5) Ensure that the quality control and quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur;</p> <p>This STANDARD is not met as evidenced by: Based on the lack of Quality Assessment Records (QAR), surveyor review of the Procedure Manual and interview with the Technical Consultants (TC), the Laboratory Director (LD) failed to ensure that the Quality Assessment (QA) program was maintained to assure the quality of laboratory services provided from 8/3/23 to 4/15/25. The findings include: 1. The laboratory did not follow the Quality Management System Annual Assessment procedure. 2. There were no QAR available for review. 3. The TC confirmed on 4/15/25 at 10:45 am, the LD did not ensure the QA program was maintained.</p>