

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 31D2311748	(X3) Date Survey Completed 04/13/2026
Name of Provider or Supplier Hackensack Meridian Urgent Care Plus Pc	Street Address, City, State 327 Franklin Avenue, Wyckoff, NJ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review survey was performed on April 13, 2026, the laboratory was found not in compliance with the following CONDITION LEVEL DEFICIENCIES D2016 - 42 C.F.R. 493.803 Condition: Successful participation proficiency testing. D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on an office review of the CASPER reports 153 and 155 and proficiency testing provider reports, the laboratory failed to achieve 80% or more for Hematology in two consecutive events for the analytes Red Blood Cell Count (RBC) and Hematocrit</p>

	<p>(HCT) with the American Association of Bioanalysts Medical Laboratory Evaluation (AAB-MLE). Refer to D2130.</p>
D2130	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>(f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the CASPER 155 report and graded results from American Association of Bioanalysts Medical Laboratory Evaluation (AAB-MLE). The laboratory failed to achieve satisfactory performance (80% or greater) for two consecutive events in for Hematology for the analytes Red Blood Cell Count (RBC), and Hemoatocrite (HCT) resulting in initial unsuccessful performance. The findings include: 1) A review of the CASPER 155 report revealed the following. a) The laboratory scored 0% for HCT in event 1-2026. b) The laboratory scored 0% for RBC in event 1-2026. c) The laboratory scored 60% for HCT in event 3-2025. d) The laboratory scored 60% for RBC in event 3-2025 2. A review of AAB-MLE graded results confirmed the laboratory failed two out of three events Proficiency Testing (PT) events.</p>
D6000	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on review of the CASPER 155 report and graded results from the American Association of Bioanalysts Medical Laboratory Evaluation (AAB-MLE). The Laboratory Director (LD) failed to provide overall management and direction to laboratory personnel to ensure that the Proficiency Testing (PT) surveys are performed satisfactorily and in compliance with Clinical Laboratory Improvement Amendments (CLIA) regulations. The findings include: 1. The LD failed to ensure PT surveys are performed satisfactorily and in compliance with CLIA regulations. Refer to D6016.</p>
D6016	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>(e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on a review of the CASPER 155 report and graded results from American Association of Bioanalysts Medical Laboratory Evaluation (AAB-MLE), the Laboratory Director (LD) failed to ensure successful participation in a Department of Health and Human Services (DHHS) approved Proficiency Testing (PT) program for</p>

two consecutive PT events for the analytes Red Blood Cell Count (RBC), and Hematocrit (HCT), resulting in initial unsuccessful performance. Refer to D2130.