

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  32D0534672	<b>(X3) Date Survey Completed</b>  05/02/2019
<b>Name of Provider or Supplier</b>  Albuquerque Dermatology Associates &	<b>Street Address, City, State</b>  5310 Homestead Rd Ne, Ste 301, Albuquerque, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The following deficiency was cited as the result of a recertification survey on 05/02/19 for 42 CFR part 493 Laboratory Requirements.
<b>D6127</b>	<p><b>TECHNICAL SUPERVISOR RESPONSIBILITIES</b> CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of personnel records and interviews with laboratory staff, the Technical Supervisor failed to perform a six-month evaluation of competency for 1 of 1 testing persons performing gross examinations of tissue samples. Findings are: A. During interview on 05/02/19 at 10:15 am, both the testing person (TP #1) and the Laboratory Supervisor stated that they did not think that a six-month competency was performed for TP #1 by the Technical Supervisor of the laboratory. TP #1 further stated that she had been employed by the laboratory since June 2018. B. No documentation of a competency evaluation was provided by the laboratory during the survey.</p>