Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	32D0933526	12/10/2025
Name of Provider or Supplier	**Street Address, City, State**	
Epiphany Dermatology Of New Mexico Llc	141 N Roadrunner Pkwy Ste 228, Las Cruces, NM	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An onsite recertification survey conducted on December 10, 2025, at Epiphany Dermatology of New Mexico LLC found the laboratory to be not in compliance with the CLIA regulations found at 42 CFR, Part 493 Laboratory Requirements, with standard deficiencies cited.
D6128	TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9) (b)(9) Thereafter, evaluations must be performed at least annually unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individuals performance must be reevaluated to include the use of the new test methodology or instrumentation. This STANDARD is not met as evidenced by: D6128: TS Competency Evaluations Based on review of the Center for Medicare and Medicaid (CMS) 209 Personnel form and personnel competency records, the Technical Supervisor failed to assess competency for 1 of 2 testing personnel (TP) performing MOHs testing in 2024 and 2025. Findings include: 1. A review of the CMS 209 personnel form revealed 2 TP performing MOHs testing. 2. A review of personnel competency records indicated no competency assessment for MOHs testing was performed for TP1 in 2024 and 2025. 3. The laboratory was asked to provide MOHs competency records for TP1 for testing years 2024 and 2025, none was provided. 4. The laboratory reports 4500 MOHs tests annually.