

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  32D0942313	<b>(X3) Date Survey Completed</b>  02/21/2018
<b>Name of Provider or Supplier</b>  Mimbres Internal Medicine	<b>Street Address, City, State</b>  122 South Gold Suite 3, Deming, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The New Mexico Department of Health conducted an offsite revisit February 20 - February 21, 2018. Based on documentation submitted by the laboratory in the Allegation of Compliance, Plan of Correction, and supporting documentation, the laboratory was found in substantial compliance with 42 CFR part 493, Laboratory requirements, for the specialty hematology. At the time of the revisit, the laboratory had been unable to resolve issues with the chemistry analyzers and studies were incomplete for rheumatoid factor.