

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 32D2022272	(X3) Date Survey Completed 01/31/2024
Name of Provider or Supplier Digestive Disease Consultants	Street Address, City, State 4381 East Lohman Ave Ste B, Las Cruces, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An onsite recertification survey conducted on January 31, 2024, at Digestive Disease Consultants found the laboratory to be in compliance with the CLIA regulations found at 42 CFR, Part 493 Laboratory Requirements, with standard deficiencies cited.
D6128	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least annually after the first year, unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's CMS form 209, competency records, and staff interview, the technical supervisor failed to evaluate annual competencies for 1 of 1 testing personnel in 2022 and 2023. Findings included: 1. Review of the laboratory's CMS form 209 revealed testing person (TP) #1, as listed on the CMS for 209, performed high complexity testing. 2. Review of TP #1's 2022 and 2023 Grossing competency records listed Quality Personnel as the competency evaluator. The Technical Supervisor failed to evaluate the annual competency. 3. In an interview on 01/31/2024 at 09:45 am, after review of the above records, testing personnel number 1 confirmed the findings.</p>