

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 32D2095249	(X3) Date Survey Completed 03/06/2024
Name of Provider or Supplier Agape Pain Management & Anesthesia Llc	Street Address, City, State 2170 E Lohman Ave Suite C, Las Cruces, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An onsite recertification survey conducted on March 6, 2024, at Agape Pain Management & Anesthesia LLC found the laboratory to be in compliance with the CLIA regulations found at 42 CFR, Part 493 Laboratory Requirements, with standard deficiencies cited.
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of ImmTox 270 Maintenance Guide, ImmTox270 Periodic Maintenance Log, and interview, the laboratory failed to follow procedures for performing required weekly maintenance on the ImmTox270 for 6 of 18 weeks from November 2023 through February 2024. Findings Included: 1. Review of the ImmTox 270 Maintenance Guide lists the following required maintenance to be done weekly: - ImmTox270 and LIS Backup - Special wash (probes and cells) - Probe Wash 2. Review of the ImmTox270 Periodic Maintenance Log revealed the following weekly maintenance being performed: - ImmTox270 and LIS Backup - Special wash (probes and cells) - Probe Wash - Clean Water Reservoir -Clean Waste Reservoir 3. Review of ImmTox270 Periodic Maintenance Log also revealed the following weeks where weekly maintenance was not performed. 1. November 13th - 17th 2023 2.December 18th - 27th 2023 3. December 25th - 29th 2023 4. January 1st - 5th 2024 5. January 15th - 19th 2024 6. February 5th - 9th 2024 4. Interview on March 6th, 2024 at 2:30 pm with the general supervisor confirmed the findings.</p>
D6065	TESTING PERSONNEL QUALIFICATIONS

CFR(s): 493.1423(b)(1)(2)(3)(4)(i)

(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located or have earned a doctoral, master's, or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; or (b)(2) Have earned an associate degree in a chemical, physical or biological science or medical laboratory technology from an accredited institution; or (b)(3) Be a high school graduate or equivalent and have successfully completed an official military medical laboratory procedures course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(4)(i) Have earned a high school diploma or equivalent; and

This STANDARD is not met as evidenced by:

Based on review of CMS 209 form and interview, the laboratory failed to provide education records for 1 (TP #4) of 7 testing personnel (TP). Findings include: 1. Review of CMS 209 form listed 7 testing personnel. 2. Education records were requested for testing personnel #4. None were provided. 3. Interview on March 6th, 2024 at 10:30 am with the general supervisor confirmed the findings.