

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 32D2123056	(X3) Date Survey Completed 10/10/2023
Name of Provider or Supplier Epiphany Dermatology Of New Mexico, Llc	Street Address, City, State 5600 Eubank Bld Ne, # 110, Albuquerque, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was performed on October 10, 2023. The laboratory was found to be in compliance with the CLIA regulations found at 42 CFR, with a standard level deficiency cited.
D6128	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least annually after the first year, unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on a review of CMS form 209, laboratory records, and confirmed in staff interview, the Technical Supervisor failed to ensure annual competency assessments were performed by personnel designated as a Technical Supervisor for 1 of 3 Testing Personnel. Findings included: 1. A review of the submitted CMS form 209 submitted by the laboratory revealed Testing Person #1 performed Mohs, Histopathology testing. 2. A review of laboratory competency records from 2022 through 2023 revealed the following: a) Testing Person #1 2022 annual assessment. Performed on 05/27/2022 by Clinical Consultant #1 (as listed on the submitted CMS form 209). b) Testing Person #1 2023 annual assessment. Performed on 05/22/2023 by Clinical Consultant #1 (as listed on the submitted CMS form 209). A designated technical supervisor failed to perform competency assessments in 2022 and 2023 for Testing Person #1. 3. During an interview on October 10, 2023 at 0940 hours in the conference room, after review of the above records, laboratory's Regional Operations Director confirmed the findings. Word key: CMS = Centers for Medicaid and Medicare Services</p>