

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 32D2188104	(X3) Date Survey Completed 11/29/2023
Name of Provider or Supplier Epiphany Dermatology Of New Mexico, Llc	Street Address, City, State No 5 Calle Medico, Ste A, Santa Fe, Santa Fe, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An onsite recertification survey conducted on 11/29/2023, at Epiphany Dermatology of New Mexico LLC found the laboratory to be in compliance with the CLIA regulations found at 42 CFR, Part 493 Laboratory Requirements, with standard deficiencies cited.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's Annual Test Volume & Proficiency Testing Programs Worksheet, Mycology and Parasitology log, KOH Quiz Score Summation, and confirmed in staff interview, the facility failed to perform twice annual accuracy of Potassium Hydroxide (KOH) Testing for 2 of 5 events from January 2021 through June 2023. Findings included: 1. Record review of the Annual Test Volume & Proficiency Testing Programs Worksheet, showed proficiency testing is completed by laboratory online quizzes and graded for accuracy. 2. Record review of KOH Quiz Score Summation showed no quiz scores for event 3 January through June 2022 and event 4 July through December 2022. The facility asked to provide scores, none were provided. 3. Review of Mycology and Parasitology Log showed the facility performed 5 KOH patient tests in 2022. 4. Interview with Mohs tech on 11/29/2023 at 10:03 am confirmed the findings.</p>