

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 32D2188104	<b>(X3) Date Survey Completed</b> 12/19/2024
<b>Name of Provider or Supplier</b> Epiphany Dermatology Of New Mexico, Llc	<b>Street Address, City, State</b> No 5 Calle Medico, Ste A, Santa Fe, Santa Fe, NM	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An onsite recertification survey conducted on December 19, 2024, at Epiphany Dermatology of New Mexico, LLC found the laboratory to be not in compliance with the CLIA regulations found at 42 CFR, Part 493 Laboratory Requirements, with standard deficiencies cited.
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Laboratory: Quality Assessment Plan, quality control logs, and interview with the manager, the laboratory failed to follow its own policy for the frequency of monitoring the room temperature and humidity in the laboratory in 2024 Finding included: 1. A review of the Laboratory: Quality Assessment Plan under, Environment, Instruments, Reagents, Materials, and Supplies stated, "The testing environment (room temperature, humidity) must be monitored each day to ensure that it meets Epiphany Dermatology or the manufacturer's stricter specifications." 2. A review of the quality control logs from 2024 revealed the laboratory failed to record room temperature and humidity on non-Mohs (micrographic surgery) testing days. 3. In an interview on 12/19/2024 at 11:30 am with the manager they stated, room temperature and humidity is only being documented on the quality control logs and currently only on days of Mohs testing", confirming the above findings. 4. The laboratory reported performing 600 Mohs tests annually.</p>