

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0127981	(X3) Date Survey Completed 03/31/2025
Name of Provider or Supplier Manhattan Dermatology Plc	Street Address, City, State 71 Park Avenue, Suite 1a, New York, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on lack of proficiency testing (PT) records and interviews with the Laboratory Director (LD) and Testing Personnel (TP), the laboratory failed to successfully enroll and participate in a histopathology specialty PT program for 2022 and 2023. FINDINGS: 1. There was no documentation of histopathology twice year verification performance for 2022 and 2023. 2. The LD and TP confirmed the findings on March 31, 2025, at approximately 12:00 P.M.</p>
D6076	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on lack of PT documentation and confirmed by interview with the TP, the LD</p>

failed to provide overall management and direction of the laboratory services. Refer to D2000.