

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0130044	(X3) Date Survey Completed 01/08/2019
Name of Provider or Supplier Yashar Hirshaut Md Pc	Street Address, City, State 6323 7th Ave, 1st Floor, Brooklyn, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3037	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(4)</p> <p>Proficiency testing records. Retain all proficiency testing records for at least 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's available Proficiency Testing (PT) records, lack of laboratory documents, and confirmed in an interview with the testing person, the laboratory failed to have documentation of American Proficiency Institute (API) PT records for the second and third events of 2018. Findings: The practice relocated from Manhattan to Brooklyn in December of 2018. On the date of this onsite survey, at approximately 3:30 PM, the PT records for the second and third events of 2018 could not be found.</p>