

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0132601	(X3) Date Survey Completed 02/05/2019
Name of Provider or Supplier Optum Medical Care Pc	Street Address, City, State 1111 Park Avenue, New York, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2006	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)</p> <p>The laboratory must examine or test, as applicable, the proficiency testing samples it receives from the proficiency testing program in the same manner as it tests patient specimens. This testing must be conducted in conformance with paragraph (b)(4) of this section. If the laboratory's patient specimen testing procedures would normally require reflex, distributive, or confirmatory testing at another laboratory, the laboratory should test the proficiency testing sample as it would a patient specimen up until the point it would refer a patient specimen to a second laboratory for any form of further testing.</p> <p>This STANDARD is not met as evidenced by: Based on a review of proficiency test records and an interview with the laboratory nurse manager/testing person, the laboratory failed to perform PT testing in the same manner as patient specimen testing. Findings Include: It was confirmed by the laboratory nurse manager/testing person on February 5, 2019, at approximately 1:00 pm that PT specimens are always performed by the nurse manager/testing person, whereas patient specimens are performed by five of five testing personnel.</p>