Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0142098	(X3) Date Survey Completed 04/12/2018
Name of Provider or Supplier Dermatology Center Of Rockland Pc	**Street Address, City, State** Prel Plaza, 60 Dutch Hill Rd Ste 18, Orangeburg, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	PROCEDURE MANUAL CFR(s): 493.1251(a) A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens. This STANDARD is not met as evidenced by: Based on a lack of procedures and an interview with the laboratory director, the laboratory failed to have a procedure for equipment maintenance. Findings Include: The laboratory director confirmed on April 12, 2018, at approximately 10:30 am that the laboratory failed to have a procedure available indicating how often maintenance will be performed on the microscope.
D6094	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(5) The laboratory director must ensure that the quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur. This STANDARD is not met as evidenced by: Based on a review of procedures and an interview with the laboratory director, the laboratory director failed to ensure that the laboratory's QA program was maintained as part of the laboratory's overall quality systems program. Refer to D5401