

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0143596	(X3) Date Survey Completed 06/12/2018
Name of Provider or Supplier Schweiger Dermatology Pllc - New Hyde Park	Street Address, City, State 3003 New Hyde Park Road Suite 204, New Hyde Park, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on a surveyor's review of Cryostat temperature records and interview with the office manager, the laboratory failed to follow the Cryostat manufacturer's temperature requirement and failed to follow the laboratory's temperature policy for Cryostat from January 2018 through the survey date. FINDINGS: 1. The laboratory performs Mohs surgery approximately once per month. The manufacturer of the Cryostat used for Mohs slide processing and the laboratory's temperature policy require that the Cryostat temperature to be monitored and documented on each day of surgery. 2. On June 12, 2018 at approximately 10:30 AM the office manager confirmed surveyor findings that the laboratory failed to monitor and document the cryostat temperature for six days from January 2018 through the date of this survey. 3. Approximately 30 patient specimens were tested and reported for Mohs surgery when the cryostat temperature was not monitored.</p>