Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 33D0163994	**(X3) Date Survey Completed** 06/16/2022
Name of Provider or Supplier Edward K Schneider Md	**Street Address, City, State** 2616 South Avenue, Wappingers Falls, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D3031	RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3) Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years. This STANDARD is not met as evidenced by: Based on the lack of Bacteriology/Throat Culture Quality Control (QC) records for the Select Strep Agar (SSA) media, 0.04 bacitracin disc and interview with the laboratory director, the laboratory failed to retain the Bacteriology/Throat Culture QC records for 2020 and 2021. FINDINGS: 1. The laboratory's Bacteriology/Throat Culture procedure requires physical characteristics, sterility & Clinical Laboratory Standards Institute (CLSI) antimicrobial performance sheets for Select Strep Agar (SSA) media and positive & negative control organism to test the reactivity the 0.04 bacitracin with each new lot and/or shipment of media & disc. 2. The laboratory's Bacteriology/Throat Culture QC records were not available for review for the calendar years 2020 and 2021. a. Current QC records for 2022 were reviewed and recorded on new QC forms for both the SSA media and 0.04 bacitracin disc. 2. The laboratory director confirmed at survey, "That the previous testing person retired and moved them to a different location. And that when preparing for the survey they could not locate them."