

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0672331	(X3) Date Survey Completed 03/27/2018
Name of Provider or Supplier Optum Medical Care Pc	Street Address, City, State 571 Chestnut Street, Cedarhurst, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6063	<p>LABORATORY TESTING PERSONNEL CFR(s): 493.1421</p> <p>The laboratory must have a sufficient number of individuals who meet the qualification requirements of 493.1423, to perform the functions specified in 493.1425 for the volume and complexity of tests performed.</p> <p>This CONDITION is not met as evidenced by: Based on a surveyor review of personnel records and confirmed in an interview with the laboratory administrator and the testing person, the laboratory director failed to ensure that one of twelve testing personnel performing moderate complexity testing met the minimum educational requirements of a high school diploma and/or had foreign education diploma evaluated prior to performing patient testing. Refer to D6065</p>
D6065	<p>TESTING PERSONNEL QUALIFICATIONS CFR(s): 493.1423(b)(1)(2)(3)(4)(i)</p> <p>(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located or have earned a doctoral, master's, or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; or (b)(2) Have earned an associate degree in a chemical, physical or biological science or medical laboratory technology from an accredited institution; or (b)(3) Be a high school graduate or equivalent and have successfully completed an official military medical laboratory procedures course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(4)(i) Have earned a high school diploma or equivalent; and</p>

This STANDARD is not met as evidenced by:
Based on a surveyor's review of personnel records and confirmed in an interview with the laboratory administrator and the laboratory testing person at the time of the survey, the laboratory director failed to ensure that one of twelve testing personnel performing moderate complexity testing met the minimum educational requirements of a high school diploma and/or had foreign education diploma evaluated prior to performing patient testing.