

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0674937	(X3) Date Survey Completed 05/28/2025
Name of Provider or Supplier Kids Care Pediatric Associates Pc	Street Address, City, State 2266 Dutch Broadway, Elmont, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5411	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(a)</p> <p>(a) Test systems must be selected by the laboratory. The testing must be performed following the manufacturer's instructions and in a manner that provides test results within the laboratory's stated performance specifications for each test system as determined under 493.1253.</p> <p>This STANDARD is not met as evidenced by: Based on review of analyzer manufacturer's requirements, laboratory Standard Operating Procedures (SOPs), as well as interview with Office Manager (OM), the laboratory failed to comply with analyzer manufacturer's requirements for humidity monitoring. FINDINGS: 1. Medonic M Series analyzer manufacturer's requirements specify humidity range of 10 - 90%. 2. There was no documentation of humidity monitoring from 2023 through survey date. 3. The current, approved SOPs did not include instructions for performing such activity. 4. OM confirmed the findings on May 28, 2025, at approximately 12:30 P.M.</p>