

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0689665	(X3) Date Survey Completed 10/16/2019
Name of Provider or Supplier Long Island Pulmonary Associates Pc	Street Address, City, State 520 Franklin Ave Suite 153, Garden City, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a proficiency testing (PT) desk review of the Center for Medicare and Medicaid Services (CMS) PT reports and PT records from the American Proficiency Institute (API) PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Cholesterol. The following scores were assigned: 2019 first event = 60% 2019 second event = 100% 2019 third event = 0% [non-participation] This is considered unsuccessful PT performance. Refer to D2096.
D2093	ROUTINE CHEMISTRY

	CFR(s): 493.841(d) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event. This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT reports and PT records from the API PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the specialty Chemistry and the test analytes Amylase, HDL Cholesterol, Total Iron (T Fe), Magnesium (Mg), Triglyceride and Uric Acid. The following scores were assigned: 2019 third event = 0% [non-participation] This is considered unsatisfactory PT performance.
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT reports and PT records from the API PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Cholesterol. The following scores were assigned: 2019 first event = 60% 2019 second event = 100% 2019 third event = 0% [non-participation] This is considered unsuccessful PT performance.
D2104	ENDOCRINOLOGY CFR(s): 493.843(d) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event. This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT reports and PT records from the API PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the sub-specialty Endocrinology and the test analytes T3-uptake, Thyroid Stimulating Hormone (TSH), Thyroxine (T4) and Triiodothyronine (T3). The following scores were assigned: 2019 third event = 0% [non-participation] This is considered unsatisfactory PT performance.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.

	This CONDITION is not met as evidenced by: Based on PT desk review of the PT CMS data reports and API PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, for the speciality Chemistry and sub-speciality Endocrinology and the test analytes Total Cholesterol, Amylase, HDL Cholesterol, T Fe, Mg, Triglyceride, Uric Acid and T3-uptake, TSH, T4 and T3. Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on PT desk review of the PT CMS data reports and API PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, for the specialty Chemistry and sub-specialty Endocrinology and the test analytes Total Cholesterol, Amylase, HDL Cholesterol, T Fe, Mg, Triglyceride, Uric Acid and T3-uptake, TSH, T4 and T3. The following scores were assigned: Total Cholesterol. 2019 second event = 60% 2019 second event = 100% 2019 third event = 0% [non-participation] This is considered unsuccessful PT performance. Speciality Chemistry and Amylase, HDL Cholesterol, T Fe, Mg, Triglyceride and Uric Acid. 2019 third event = 0% [non-participation] Specialty Endocrinology and the test analytes T3-uptake, TSH, T4 and T3. The following scores were assigned: 2019 third event = 0% [non-participation] This is considered unsatisfactory PT performance.