

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 33D0692168 | (X3) Date Survey Completed 01/31/2019 |
| Name of Provider or Supplier Allied Physicians Group Plc | Street Address, City, State 2611 Corporal Kennedy Street, Bayside, NY | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D5209 | <p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on a review of competency procedures, records and an interview with the laboratory director and office manager, the laboratory failed to have a complete policy /procedure to assess employees competency. Finding Include: At approximately 2:00 pm on January 31, 2019, it was confirmed by the laboratory director and office manager that the laboratory failed to include in their competency procedure direct observation of routine patient test performance (including patient preparation, if applicable, specimen handling, processing, and testing); and the performance of instrument maintenance and function checks.</p> |