

|                                                                                                                            |                                                                            |                                                     |
|----------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------|-----------------------------------------------------|
| <b>Statement of Deficiencies</b>                                                                                           | <b>(X1) Provider/Supplier/CLIA Identification Number</b><br><br>33D0855598 | <b>(X3) Date Survey Completed</b><br><br>08/19/2024 |
| <b>Name of Provider or Supplier</b><br><br>Oneida Pediatric Group Pc                                                       | <b>Street Address, City, State</b><br><br>421 Main Street, Oneida, NY      |                                                     |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. |                                                                            |                                                     |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |
|--------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| D0000              | Based on a proficiency testing (PT) desk review survey performed on August 19, 2024, the laboratory was found to be out of compliance based on the following CONDITION LEVEL DEFICIENCIES: D2016 - 42 C.F.R. 493.803 Condition: Successful participation. D6000 - 42 C.F.R. 493.1403 Condition: Laboratory Director, moderate complexity.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |
| D2016              | <p>SUCCESSFUL PARTICIPATION<br/>CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by:<br/>Based on review of Centers for Medicare &amp; Medicaid Services (CMS) Proficiency Testing (PT) Certification and Survey Provider Enhanced Reporting system (CASPER 0155D) and College of American Pathologists (CAP) PT summary reports,</p> |

|              |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |
|--------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|              | <p>the laboratory failed to successfully participate in the CMS approved PT program for two of three consecutive testing events in the Hematology subspecialty for the Cell ID or WBC Diff, HGB (Hemoglobin) Non-Waived, and WBC (White Blood Cell) Count test analytes in 2023 and 2024, resulting in non-initial unsuccessful performance. Refer to D2130.</p>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |
| <b>D2130</b> | <p><b>HEMATOLOGY</b><br/>CFR(s): 493.851(f)</p> <p>Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by:<br/>Based on CMS PT CASPER 0155D and CAP PT summary reports from 2023 and 2024, the laboratory failed to achieve satisfactory performance (80% or greater) for two of three consecutive testing events in the Hematology subspecialty for the Cell ID or WBC Diff, HGB (Non-Waived), and WBC Count test analytes. FINDINGS: 1. A review of the CASPER 155 report revealed the following unsatisfactory scores: Hematology Subspecialty: 2023 Third Event = 0% 2024 First Event = 70% Cell ID or WBC Diff Test Analyte: 2023 Third Event = 0% 2024 First Event = 60% HGB (Non-Waived) Test Analyte: 2023 Third Event = 0% 2024 First Event = 60% WBC Count Test Analyte: 2023 Third Event = 0% 2024 First Event = 40% 2. A review of the proficiency testing scores from CAP (2023 and 2024) confirmed the above findings.</p> |
| <b>D6000</b> | <p><b>MODERATE COMPLEXITY LABORATORY DIRECTOR</b><br/>CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by:<br/>Based on review of CMS PT CASPER 0155D and CAP PT summary reports from 2023 and 2024, the laboratory director (LD) failed to failed to provide overall management and direction of the laboratory services. Refer to D2016.</p>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| <b>D6016</b> | <p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b><br/>CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by:<br/>Based on review of CMS PT CASPER 0155D and CAP PT 2023-3 and 2024-1 summary reports, the LD failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2130.</p>                                                                                                                                                                                                                                                                         |