

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0880382	(X3) Date Survey Completed 05/15/2020
Name of Provider or Supplier Comprehensive Pediatrics Pc	Street Address, City, State 1407 West 6th Street, Brooklyn, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a proficiency testing (PT) desk review of the Center for Medicare and Medicaid Services (CMS) PT reports and PT records from the American Association of Bioanalysts (AAB) PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analytes Red Blood Cell Count (RBC), Hemoglobin (Hgb) and Hematocrit (Hct). The following scores were assigned RBC/Hct 2019 second event = 60% 2019 third event = 100% 2020 first event = 0% [non-participation] Hgb 2019 second event = 60% 2019 third event = 80% 2020 first event = 0% [non-participation] This is considered unsuccessful PT performance. Refer to D2130.

D2123	HEMATOLOGY CFR(s): 493.851(c) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events. This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT reports and PT records from the AAB PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the specialty Hematology and the test analytes White blood Cell count (WBC), Platelets and WBC Differential. The following scores were assigned 2020 first event = 0% [non-participation] This is considered unsatisfactory PT performance.
D2130	HEMATOLOGY CFR(s): 493.851(f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT reports and PT records from the AAB PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analytes RBC, Hgb and Hct. The following scores were assigned RBC/Hct 2019 second event = 60% 2019 third event = 100% 2020 first event = 0% [non-participation] Hgb 2019 second event = 60% 2019 third event = 80% 2020 first event = 0% [non-participation] This is considered unsuccessful PT performance.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on PT desk review of the PT CMS data reports and AAB PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, for the specialty Hematology and the test analytes WBC, RBC, Hgb, Hct, Cell Differential and Platelets. Refer to D6016.

D6016

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1407(e)(4)(i)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:

Based on PT desk review of the PT CMS data reports and AAB PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, for the specialty Hematology and the test analytes WBC, RBC, Hgb, Hct, WBC Differential and Platelets. The following scores were assigned RBC/Hct 2019 second event = 60% 2019 third event = 100% 2020 first event = 0% [non-participation] Hgb 2019 second event = 60% 2019 third event = 80% 2020 first event = 0% [non-participation] This is considered unsuccessful PT performance. Specialty Hematology and analytes WBC, WBC differential and Platelets 2020 first event = 0% [non participation] This is considered unsatisfactory PT performance.