

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0882284	(X3) Date Survey Completed 07/05/2023
Name of Provider or Supplier Murray Hill Urology, Pc	Street Address, City, State 120 East 34 Street, New York, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2004	<p>ENROLLMENT CFR(s): 493.801(a)(3)</p> <p>For each specialty, subspecialty and analyte or test, participate in one approved proficiency testing program or programs, for one year before designating a different program and must notify CMS before any change in designation;</p> <p>This STANDARD is not met as evidenced by: Based on Proficiency Test (PT) desk review of the Center for Medicaid & Medicare Services (CMS) PT data reports and PT summary reports from American Proficiency Institute (API) and College of American Pathologist (CAP) PT programs, the laboratory failed to participate in one approved proficiency testing program for one year before designating a different program. FINDINGS: The laboratory is enrolled in both API and CAP PT programs for General Immunology, Chemistry, Endocrinology testing for the same test analyte's and reporting both results to CMS for 2023 1st & 2nd test events.</p>
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical</p>

assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.

This CONDITION is not met as evidenced by:

Based on PT desk review of the CMS PT data reports and PT summary reports from API PT program, the laboratory failed to participate successfully in CMS approved proficiency testing program, for the test analyte Chloride. The following scores were assigned: 2022 second event = 20% 2022 third event = 0% 2023 first event = 100% 2023 second event = 40% This is considered unsuccessful PT performance. Refer to D2096

D2087

ROUTINE CHEMISTRY

CFR(s): 493.841(a)

Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.

This STANDARD is not met as evidenced by:

Based on a PT desk review of the CMS PT reports and PT records from the API PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Calcium. The following scores were assigned: 2023 second event = 0% This is considered a unsatisfactory PT performance.

D2096

ROUTINE CHEMISTRY

CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on PT desk review of the CMS PT data reports and PT summary reports from API PT program, the laboratory failed to participate successfully in CMS approved proficiency testing program, for the test analyte Chloride. The following scores were assigned: 2022 second event = 20% 2022 third event = 0% 2023 first event = 100% 2023 second event = 40% This is considered unsuccessful PT performance.

D6000

MODERATE COMPLEXITY LABORATORY DIRECTOR

CFR(s): 493.1403

The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.

This CONDITION is not met as evidenced by:
Based on PT desk review of the CMS PT and API PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the test analyte's Chloride and Total Calcium Refer D6016

D6016

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1407(e)(4)(i)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:
Based on PT desk review of the CMS PT and API PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the test analyte's Chloride and Total Calcium. The following scores were assigned: Chloride 2022 second event = 20% 2022 third event = 0% 2023 first event = 100% 2023 second event = 40% This is considered unsuccessful PT performance. Total Calcium 2023 second event = 0% This is considered a unsatisfactory PT performance.