

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0912256	(X3) Date Survey Completed 06/13/2019
Name of Provider or Supplier Itchie Statfeld Md Pllc	Street Address, City, State 531 Bedford Avenue, Brooklyn, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: . Based on the surveyor's review of American Proficiency Institute (API) Proficiency Testing (PT) records and an interview with the laboratory director, the laboratory failed to successfully participate in a PT program approved by the Centers for Medicare and Medicaid Services (CMS) for the test analyte Platelets for the first event of 2019. The following score was assigned: Platelets: 2019 first event = 60% This is considered unsatisfactory PT performance.</p>