

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0934371	(X3) Date Survey Completed 01/18/2018
Name of Provider or Supplier Broadway Pediatrics Md Pc	Street Address, City, State 4250 Broadway, Suite 1 C, New York, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of Proficiency Test (PT) records from the College of American Pathology (CAP) and confirmed in an interview with the laboratory director at the time of the survey, the laboratory failed to perform remediation for the less than 100% hematology PT score for the 2017 1st event for RBC = 80 %.</p>
D5403	<p>PROCEDURE MANUAL CFR(s): 493.1251(b)</p> <p>The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals (normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the</p>

protocol for reporting imminently life threatening results, or panic, or alert values.
(14) Description of the course of action to take if a test system becomes inoperable.

This STANDARD is not met as evidenced by:

Based on a review of the laboratory procedure manual and confirmed by the laboratory director in an interview on January 18, 2018 at approximately 1:15 pm, the laboratory failed to have a complete procedure manual to include a procedure for lot to lot verification of new control material used on the Horiba Micros 60 hematology analyzer and a quality control acceptable criteria procedure.

D6021

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1407(e)(5)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(5) Ensure that quality assessment programs are established and maintained to assure the quality of laboratory services provided.

This STANDARD is not met as evidenced by:

Based on a review of records and confirmed in an interview with the laboratory director on January 18, 2018 at approximately 1:15 PM, the laboratory director failed to ensure that the QA program for hematology testing was maintained to ensure quality laboratory services. Refer to D5403 and 5211.