

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D0961627	(X3) Date Survey Completed 04/15/2025
Name of Provider or Supplier Citywide Dermatology Pllc	Street Address, City, State 3010 Westchester Avenue - Ste 207, Purchase, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on review of Standard Operating Procedures (SOPs), lack of Proficiency Testing (PT) records, as well as interview with the Practice Manager (PM), the laboratory failed to successfully participate in a histopathology specialty PT program. FINDINGS: 1. There was no documentation of Mohs histopathology twice year verification performance for 2024. 2. The current, approved SOPs did not include instructions for performing such activity. 3. The PM confirmed the findings on April 15, 2025, at approximately 12:30 P.M.</p>
D6076	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by:</p>

Based on review of SOPs, lack of PT documentation, as well as interview with the PM, the Laboratory Director (LD) failed to provide overall management and direction of the laboratory services. Refer to D2000.