

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1000123	(X3) Date Survey Completed 05/09/2023
Name of Provider or Supplier Naomi Chaim-Watman Md Pllc	Street Address, City, State 303 East Park Ave, Suite A, Long Beach, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5203	<p>SPECIMEN IDENTIFICATION AND INTEGRITY CFR(s): 493.1232</p> <p>The laboratory must establish and follow written policies and procedures that ensure positive identification and optimum integrity of a patient's specimen from the time of collection or receipt of the specimen through completion of testing and reporting of results.</p> <p>This STANDARD is not met as evidenced by: Based on the review of laboratory's Standard Operating "Specimen Collection," the procedure failed to include second identifier unique to the patient. Confirmed finding on an interview with laboratory director on 5/9/2023 about 10:45am.</p>
D5311	<p>SPECIMEN SUBMISSION, HANDLING, AND REFERRAL CFR(s): 493.1242(a)</p> <p>The laboratory must establish and follow written policies and procedures for each of the following, if applicable: (1) Patient preparation. (2) Specimen collection. (3) Specimen labeling, including patient name or unique patient identifier and, when appropriate, specimen source. (4) Specimen storage and preservation. (5) Conditions for specimen transportation. (6) Specimen processing. (7) Specimen acceptability and rejection. (8) Specimen referral.</p> <p>This STANDARD is not met as evidenced by: Based on the review of laboratory's Standard Operating procedure, the laboratory failed to establish written procedure of phlebotomy/blood collection of specimens to be sent to reference laboratory. Confirmed finding on an interview with laboratory director on 5/9/2023 about 11:15am.</p>