

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1002531	(X3) Date Survey Completed 06/23/2021
Name of Provider or Supplier Women's Health Care Pc	Street Address, City, State 87-16 Avon Street, Jamaica Est, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a proficiency testing (PT) desk review of the Center for Medicare and Medicaid Services (CMS) PT data reports, and College of American Proficiencies (CAP) PT records, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Bilirubin (T. Bili). The following scores were assigned: 2020 second event = 20% 2020 third event = 100% 2021 first event = 60% This is considered an unsuccessful PT performance Refer to D2096</p>
D2087	ROUTINE CHEMISTRY

	<p>CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on PT desk review of the CMS PT data reports and PT records from the CAP PT program, the laboratory failed to participate successfully in proficiency testing for the test analyte's Alanine Aminotransferase (ALT/SGPT), Alkaline Phosphates, Albumin, Aspartate Aminotransferase (AST/SGOT), Chloride (CL), Total Calcium (T. Ca), Total Protein (T. Protein) Creatine, Glucose, Magnesium (Mg), Potassium (K), Sodium (Na), Blood Urea Nitrogen (BUN) and Uric Acid The following score was assigned: ALT/SGPT, Alkaline Phosphates, Albumin, AST/SGOT, CL, Creatine, Glucose, Mg, K, Na, T. Protein, and Uric Acid 2021 first event = 60% BUN 2021 first event= 20% SGOT and T CA 2021 first event = 40% This is considered unsatisfactory PT performance.</p>
<p>D2088</p>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT data reports and PT records from the CAP PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the specialty Chemistry. The following scores were assigned: 2021 first event = 54% This is considered a unsatisfactory PT performance.</p>
<p>D2096</p>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p> <p>Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT data reports, and CAP PT program records, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Bilirubin (T. Bili). The following scores were assigned: 2020 second event = 20% 2020 third event = 100% 2021 first event = 60% This is considered an unsuccessful PT performance</p>
<p>D2098</p>	<p>ENDOCRINOLOGY CFR(s): 493.843(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p>

	<p>This STANDARD is not met as evidenced by: Based on PT desk review of the CMS PT data reports and CAP PT program records, the laboratory failed to participate successfully in proficiency testing for the test analyte Triiodothyronne (T3) The following score was assigned: 2020 second event = 0% This is considered unsatisfactory PT performance.</p>
<p>D6000</p>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on PT desk review of the CMS PT data reports and CAP PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the specialty Routine Chemistry and the test analyte's ALT/SGPT, Alkaline Phosphates, Albumin, AST/SGOT, CL, T. Bili., Creatine, Glucose, Mg, K, Na, T. Protein, BUN, T. CA, Uric Acid and T3 Refer to D6016.</p>
<p>D6016</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on PT desk review of the CMS PT data reports and CAP PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the specialty Routine Chemistry and the test analyte's ALT/SGPT, Alkaline Phosphates, Albumin, AST/SGOT, CL, T. Bili., Creatine, Glucose, Mg, K, Na, T. Protein, BUN, T. CA, Uric Acid and T3. The following scores were assigned: Total Bilirubin (T. Bili). 2020 second event = 20% 2020 third event = 100% 2021 first event = 60% This is considered an unsuccessful PT performance specialty Chemistry. 2021 first event = 54% ALT/SGPT, Alkaline Phosphates, Albumin, AST/SGOT, CL, Creatine, Glucose, Mg, K, Na, T. Protein, and Uric Acid 2021 first event = 60% BUN 2021 first event= 20% AST/SGOT and T CA 2021 first event = 40% Triiodothyronne (T3) 2020 second event = 0% This is considered unsatisfactory PT performance.</p>