

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 33D1012428 | (X3) Date Survey Completed 08/15/2018 |
| Name of Provider or Supplier Anatoly Belilovsky Md Pllc | Street Address, City, State 645 Bay Street, Staten Island, NY | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|---|
| D5429 | <p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's review of the Traceable Digital Thermometers packet insert , observation of the refrigerator thermometer and an interview with the technical consultant, the laboratory failed to follow the manufacturer's requirements for calibration of the Traceable thermometer used to record the laboratory refrigerator's temperatures. FINDINGS: The technical consultant confirmed on August 15, 2018 at approximately 11:00 AM, that the Traceable thermometer used to record the laboratory refrigerator's temperature was out of calibration from 2/19/18 through survey date.</p> |