

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1019019	(X3) Date Survey Completed 03/22/2023
Name of Provider or Supplier Boro Park Obgyn, Pc	Street Address, City, State 5925 15th Avenue, Brooklyn, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6115	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(2)</p> <p>The technical supervisor is responsible for verification of the test procedures performed and establishment of the laboratory's test performance characteristics, including the precision and accuracy of each test and test system.</p> <p>This STANDARD is not met as evidenced by:</p> <p>A. Based on the microscopic review of 396 negative gynecologic cytology cases /slides from February 2023 and confirmation by the Technical Supervisor on March 21, 2023 and April 4, 2023, the Technical Supervisor failed to verify the accuracy of three gynecologic cytology tests. C2023-1648 02/05/2023 FocalPoint Guided Imaging LABORATORY DIAGNOSIS: Negative for Intraepithelial Lesion or Malignancy SURVEY TEAM DIAGNOSIS: Unsatisfactory for Interpretation. Insufficient Cellularity TECHNICAL SUPERVISOR DIAGNOSIS: Unsatisfactory C2023-1789 02/11/2023 FocalPoint Guided Imaging LABORATORY DIAGNOSIS: Negative for Intraepithelial Lesion or Malignancy SURVEY TEAM DIAGNOSIS: Unsatisfactory for Interpretation. Insufficient Cellularity TECHNICAL SUPERVISOR DIAGNOSIS: Unsatisfactory. Scant C2023-2388 02/19/2023 FocalPoint Guided Imaging LABORATORY DIAGNOSIS: Negative for Intraepithelial Lesion or Malignancy SURVEY TEAM DIAGNOSIS: Unsatisfactory for Interpretation. Insufficient Cellularity TECHNICAL SUPERVISOR DIAGNOSIS: Unsatisfactory B.</p> <p>B. Based on the microscopic review of 46 non- negative gynecologic cytology cases /slides from January and February 2023 and confirmation by the Technical Supervisor on March 21, 2023, the Technical Supervisor failed to verify the accuracy of one gynecologic cytology test. C2023-0743 02/22/2023 FocalPoint Guided Imaging LABORATORY DIAGNOSIS: Low Grade Squamous Intraepithelial Lesion SURVEY TEAM DIAGNOSIS: Negative for Intraepithelial Lesion or Malignancy TECHNICAL SUPERVISOR DIAGNOSIS: Negative for Intraepithelial Lesion or Malignancy</p>

D9999

By agreement between ASCT Services, Inc. and CMS, information provided for CMS's completion of CMS Form 670 are ASCT Services, Inc. averages only. This information is confidential and proprietary to ASCT Services, Inc., is exempt under the Freedom of Information Act (5 U.S.C. 552 et seq.), and shall be used for federal government purposes only.