

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1027412	(X3) Date Survey Completed 07/05/2023
Name of Provider or Supplier Gregory Shifrin Md Pc Ob Gyn	Street Address, City, State 1766 East 12th Street, Basement, Brooklyn, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on Proficiency Test (PT) desk review of the Center for Medicaid & Medicare Services (CMS) PT data reports and PT summary reports from American Proficiency Institute (API) PT program, the laboratory failed to participate successfully in CMS approved proficiency testing program, for the speciality General Immunology and the test analyte Alpha Fetoprotein (AFP). The following scores were assigned: Speciality General Immunology and AFP 2023 first event = 40% 2023 second event = 0% This is considered unsuccessful PT performance. Refer to D2084 and D2085
D2084	GENERAL IMMUNOLOGY

	CFR(s): 493.837(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on PT desk review of the CMS PT data reports and PT summary reports from API PT program, the laboratory failed to participate successfully in CMS approved proficiency testing program, for the test analyte Alpha Fetoprotein (AFP). The following scores were assigned: AFP 2023 first event = 40% 2023 second event = 0% This is considered unsuccessful PT performance.
D2085	GENERAL IMMUNOLOGY CFR(s): 493.837(g) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on PT desk review of the CMS PT data reports and PT summary reports from API PT program, the laboratory failed to participate successfully in CMS approved proficiency testing program, for the speciality General Immunology. The following scores were assigned: Speciality General Immunology 2023 first event = 40% 2023 second event = 0% This is considered unsuccessful PT performance.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on PT desk review of the CMS PT and API PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the speciality for the speciality for the speciality General Immunology and the test analyte AFP. Refer to D6016
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:

Based on PT desk review of the CMS PT and API PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the speciality for the speciality for the speciality General Immunology and the test analyte AFP. The following scores were assigned: Speciality General Immunology and AFP 2023 first event = 40% 2023 second event = 0% This is considered unsuccessful PT performance.