

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1036201	(X3) Date Survey Completed 05/19/2026
Name of Provider or Supplier Medical Pain Consultant	Street Address, City, State 2127 Dryden Road, Freeville, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5435	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(b)(2)</p> <p>(b)(2)(i) Define a function check protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting. (b)(2)(ii) Perform and document the function checks, including background or baseline checks, specified in paragraph (b)(2)(i) of this section. Function checks must be within the laboratory's established limits before patient testing is conducted.</p> <p>This STANDARD is not met as evidenced by: Based on lack of Medical Pain Consultants, Osteopathy of Tompkins County, PC Standard Operating Procedures (SOPs), thermometer calibration records, as well as interview with the Testing Person (TP), the laboratory failed to define a function check protocol that ensures instrument performance that is necessary for accurate and reliable test results and result reporting. FINDINGS: 1. There was no documentation of calibration for the Acurite and Ithaca Guitar Works promotion branded digital thermometer hygrometers utilized for monitoring ambient room temperature and humidity in the laboratory. No serial or model number was available for both the respective Acurite and Ithaca Guitar Works thermometer hygrometers. 2. The Traceable Thomas Scientific digital thermometer, serial number (SN) 221352774, model 1222W14, utilized for monitoring the temperature of the Frigidaire refrigerator where reagents, calibrators, controls, proficiency testing samples were stored, featured a tag indicating recalibration due March 15, 2024. No certificate of calibration or documentation was available. 3. The Traceable Thomas Scientific digital thermometer, SN: 221352768, model 1222W14, utilized for monitoring the temperature of the Frigidaire freezer where patient urine specimens were stored, featured a tag indicating recalibration due March 15, 2024. It was noted that the respective calibration thermometer certificate was retained. 4. Medical Pain Consultants, Osteopathy of Tompkins County, PC laboratory failed to draft and</p>

approve procedures for thermometer calibration and certificate retention. 5. The TP confirmed the findings on May 19, 2026, at approximately 1:15 P.M.