

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1043447	(X3) Date Survey Completed 09/15/2022
Name of Provider or Supplier Optum Medical Care Pc	Street Address, City, State 340 Montauk Highway, West Islip, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's competency assessment policies, competency evaluation records and an interview with the laboratory operation manager, the laboratory failed to follow their establish written policies and procedures to assess the competency of the laboratory testing personnel. FINDINGS: The laboratory operation manager confirmed on September 15, 2022 at approximately 11am, that the laboratory did not follow the established competency evaluation policy. The laboratory did not perform 2021 annual competency evaluation for three of ten laboratory testing personnel.</p>