

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1059894	(X3) Date Survey Completed 12/31/2018
Name of Provider or Supplier Advanced Gynecology And Surgery, Pc	Street Address, City, State 1711 Sheepshead Bay Rd, 1st Floor, Brooklyn, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a proficiency testing (PT) desk review of the Center for Medicare and Medicaid Services (CMS) PT reports and PT records from the College of American Pathologists (CAP) PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Thyroxine (T4). The following scores were assigned: 2018 second event = 0% 2018 third event = 0% This is considered unsuccessful PT performance [non-participation]. Refer to D2107.
D2107	ENDOCRINOLOGY

	CFR(s): 493.843(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on a PT desk review of the CMS PT reports and PT records from the CAP PT program, the laboratory failed to participate and perform successfully in a PT program, approved by CMS, for the test analyte Total Thyroxine (T4). The following scores were assigned: 2018 second event = 0% 2018 third event = 0% This is considered unsuccessful PT performance [non-participation].
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on PT desk review of the CMS PT and CAP PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the test analyte Total Thyroxine (T4). Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on PT desk review of the CMS PT and CAP PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the test analyte Total Thyroxine (T4). The following scores were assigned: 2018 second event = 0% 2018 third event = 0% This is considered unsuccessful PT performance [non-participation].