

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D1088556	(X3) Date Survey Completed 10/22/2018
Name of Provider or Supplier New York Cosmetic Skin & Laser Surgery Center	Street Address, City, State 901 Stewart Ave, Suite 240, Garden City, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on a lack of room temperature and humidity records and an interview with the laboratory director, the laboratory failed to follow the manufacturer's instructions to monitor and document the room temperature and humidity where testing is performed. Findings Include: It was confirmed by the laboratory director, on October 22, 2018 at approximately 12:30 PM that the laboratory failed to follow the manufacturer's written criteria to monitor and document the temperature and humidity of the room where Moh's testing is performed from November 2016 through the date of this survey.</p>
D6094	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(5)</p> <p>The laboratory director must ensure that the quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.</p> <p>This STANDARD is not met as evidenced by:</p>

Based on a review of the laboratory's records and an interview with the laboratory director, the laboratory director failed to ensure that the laboratory followed the manufacturer's instruction to monitor and document the temperature and humidity of the room where testing is performed. Refer to D5413