

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D2042099	(X3) Date Survey Completed 12/18/2021
Name of Provider or Supplier Nephrology Associates Of Watertown, Pc	Street Address, City, State 19316 Us Route 11, Suite B, Watertown, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on Proficiency Test (PT) desk review of the Center for Medicaid & Medicare Services (CMS) PT data reports and PT records from Medical Laboratory Evaluation Program (MLE) PT program, the laboratory failed to participate successfully in proficiency testing for the test analyte Sodium. The following scores were assigned: 2021 second event = 0% 2021 third event = 60% This is considered unsuccessful PT performance. Refer to D2096
D2094	ROUTINE CHEMISTRY CFR(s): 493.841(e)

	(1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event. This STANDARD is not met as evidenced by: Based on PT desk review of CMS PT data reports and PT records from MLE PT program, the laboratory failed to participate successfully in proficiency testing for the test analyte's Potassium (K) and Creatinine. The following scores were assigned: Potassium 2021 second event = 60% Creatinine 2021 third event = 40% This is considered unsatisfactory PT performance
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on PT desk review of CMS PT data reports and PT records from MLE PT program, the laboratory failed to participate successfully in proficiency testing for the test analyte Sodium. The following scores were assigned: 2021 second event = 0% 2021 third event = 60% This is considered unsuccessful PT performance.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on PT desk review of the CMS PT data reports and MLE PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the test analyte's Sodium, Potassium (K) and Creatinine. Refer to D6016
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory

director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:

Based on PT desk review of the CMS PT data reports and MLE PT program records, the laboratory director failed to fulfill the laboratory director's responsibilities and ensure that the laboratory achieved a satisfactory performance and successfully participate in a PT program, approved by CMS, for the test analyte's Sodium, Potassium (K) and Creatinine. The following scores were assigned: Sodium 2021 second event = 0% 2021 third event = 60% This is considered unsuccessful PT performance. Potassium 2021 second event = 60% Creatinine 2021 third event = 40% This is considered unsatisfactory PT performance.