

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D2047877	(X3) Date Survey Completed 07/14/2021
Name of Provider or Supplier As Medical Pc	Street Address, City, State 2493 Richmond Rd Suite 2, Staten Island, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5032	<p>CYTOLOGY CFR(s): 493.1221</p> <p>If the laboratory provides services in the subspecialty of Cytology, the laboratory must meet the requirements specified in 493.1230 through 493.1256, 493.1274, and 493.1281 through 493.1299.</p> <p>This CONDITION is not met as evidenced by: Based on surveyor's review of laboratory policies and procedures, laboratory records and confirmed in an interview with the pathologist/laboratory director, the laboratory failed to establish, reassess and document a workload limit for the laboratory director /technical supervisor (refer to D5633, D5637, D5645 and D5647);the pathologist failed to perform and document the quality control (QC) acceptability of the Hematoxylin and Eosin (H & E) and immunostain stain used for histopathology slides (refer to D5437, D5601). The cumulative effect of these systemic problems resulted in the laboratory's inability to ensure the accuracy and reliability of patient test results in the subspecialty of Cytology.</p>
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's review of the twice year verification records for histology and cytology slides and an interview with the pathologist/laboratory director/technical supervisor, the laboratory failed to verify the accuracy for the urine cytology slides and the Fluorescence In Situ Hybridization (FISH) images from August 3, 2020</p>

through July, 5, 2021. FINDINGS: The pathologist/laboratory director/technical supervisor confirmed on July 14, 2021 at approximately 12:30 PM that the laboratory failed to verify the accuracy for the urine cytology slides and the FISH images from August 3, 2020 through July 5, 2021. a. Approximately 150 patients were tested and reported for urine cytology. b. Approximately 50 patients were tested and reported for FISH.

D5291

GENERAL LABORATORY SYSTEMS QUALITY ASSESSMENT
CFR(s): 493.1239(a)

The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and, when indicated, correct problems identified in the general laboratory systems requirements specified at 493.1231 through 493.1236.

This STANDARD is not met as evidenced by:
Based on surveyor's review of the laboratory's Quality Assessment (QA) policy, lack of QA review records for the calendar year 2020 and an interview with the pathologist /laboratory director/technical supervisor, the laboratory failed to follow their established QA policy for an ongoing mechanism to monitor, assess and correct problems identified in the general laboratory systems. FINDINGS: 1. The pathologist /laboratory director/technical supervisor confirmed on July 14, 2021 at approximately 1:15 PM, that the laboratory failed to follow their established QA policy for an ongoing mechanism to monitor, assess and correct problems identified in the general laboratory systems. 2. The laboratory's QA policy requires an annual review of all phases of laboratory's histology and cytology testing. a. the laboratory failed to perform and document the QA review for the calendar year 2020. 3. The laboratory failed to identify and take corrective action for the following issues: a. failure to document the stain quality for the Hematoxylin and Eosin (H&E) and reactivity for the immunostain for the histology slides b. failure to document the Papanicolaou (PAP) stain for the urine cytology slides 4. The pathologist failed to document his cytology workload from August 3, 2020 through survey date. a. failure to reassess the workload every six months

D5473

CONTROL PROCEDURES
CFR(s): 493.1256(e)(2)(g)

(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate. (g) The laboratory must document all control procedures performed.

This STANDARD is not met as evidenced by:
Based on surveyor's review of the histology log sheets and lack of quality control records and an interview with the pathologist/ laboratory director/technical supervisor, the pathologist did not perform and document the quality control (QC) acceptability of the Hematoxylin and Eosin (H&E) stain used for histopathology slides. FINDINGS 1. The pathologist/laboratory director/technical supervisor confirmed on July 14, 2021 at approximately 1:30 PM, the pathologist did not perform and document the quality control (QC) acceptability of the Hematoxylin and Eosin (H & E) stain used for

	<p>histopathology slides from August 8, 2020 through July 14, 2021. a. Approximately 400 patient slides were tested, and results reported for H&E stains from August 8, 2020 through July 14, 2021.</p>
<p>D5601</p>	<p>HISTOPATHOLOGY CFR(s): 493.1273(a)(f)</p> <p>(a) As specified in 493.1256(e)(3), fluorescent and immunohistochemical stains must be checked for positive and negative reactivity each time of use. For all other differential or special stains, a control slide of known reactivity must be stained with each patient slide or group of patient slides. Reactions of the control slide with each special stain must be documented. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's review of histopathology QC records and an interview with the pathologist/laboratory director/technical supervisor at the time of this survey, the laboratory director failed to document the positive and negative reactivity and the staining characteristics of the H&E stain and immunostain's on each day of interpretation from August 3, 2020 through July 5,2021. FINDINGS: 1. The pathologist/laboratory director/technical supervisor confirmed on July 14, 2021 at approximately 12:30 PM, findings that the pathologist failed to document the positive and negative reactivity and the staining characteristics of the H&E stains on each day of interpretation from August 3, 2020 through July 5, 2021. 2. Approximately 400 patients histopathology slides were read and reported during above time frame.</p>
<p>D5631</p>	<p>CYTOLOGY CFR(s): 493.1274(c)(6)</p> <p>(c) Control procedures. The laboratory must establish and follow written policies and procedures for a program designed to detect errors in the performance of cytologic examinations and the reporting of results. The program must include the following: (c) (6) An evaluation of the case reviews of each individual examining slides against the laboratory's overall statistical values, documentation of any discrepancies, including reasons for the deviation, and, if appropriate, corrective actions taken.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's review of cytology procedures, QA laboratory records and an interview with the pathologist/laboratory director/technical supervisor, the laboratory failed to establish and follow procedure to include: an evaluation of the cases review by the individual against the laboratory's overall statistical values, documentation of any discrepancies, including reasons for the deviation, and, if appropriate, the corrective actions taken FINDINGS: The pathologist/laboratory director/technical supervisor, confirmed on July 14, 2021 at approximately 1:00 PM , that the laboratory failed to establish and follow a procedure to include: an evaluation of the cases review by the individual against the laboratory's overall statistical values, documentation of any discrepancies, including reasons for the deviation, and, if appropriate, the corrective actions taken.</p>
<p>D5633</p>	<p>CYTOLOGY CFR(s): 493.1274(d)(1)</p>

(d) Workload limits. The laboratory must establish and follow written policies and procedures that ensure the following: (d)(1) The technical supervisor establishes a maximum workload limit for each individual who performs primary screening.

This STANDARD is not met as evidenced by:

Based on surveyor's review of the pathology laboratory procedure manual, lack of laboratory documents, and an interview with the pathologist/laboratory director /technical supervisor, the laboratory failed to follow laboratory's written procedures to ensure that the pathologist ,acting as the technical supervisor, established workload limits for the pathologist when performing primary screening of the non-gynecologic specimens from August 3, 2020 through July 14, 2021. Approximately 200 non-gynecologic cytology slides were reviewed and reported during this time frame.

D5637

CYTOLOGY

CFR(s): 493.1274(d)(1)(ii)

(d) Workload limits. The laboratory must establish and follow written policies and procedures that ensure the following: (d)(1)(ii) Each individual's workload limit is reassessed at least every 6 months and adjusted when necessary.

This STANDARD is not met as evidenced by:

Based on surveyor's review of the laboratory's cytology procedure manual, laboratory records and an interview with the pathologist/laboratory director/technical supervisor, the laboratory failed to established written policies and procedures to ensure that workload limits would be reassessed at least every 6 months and adjusted when necessary for the pathologist who performs the primary screening of non-gynecologic cytology slides. FINDINGS: The pathologist/laboratory director/technical supervisor, confirmed on July 14, 2021 at approximately 1:30 PM, that the laboratory failed to established written policies and procedures to ensure that workload limits would be reassessed at least every 6 months and adjusted when necessary for the pathologist who performs the primary screening of non-gynecologic cytology slides.

D5645

CYTOLOGY

CFR(s): 493.1274(d)(3)

(d) Workload limits. The laboratory must establish and follow written policies and procedures that ensure the following: (d)(3) The laboratory must maintain records of the total number of slides examined by each individual during each 24-hour period and the number of hours spent examining slides in the 24-hour period irrespective of the site or laboratory.

This STANDARD is not met as evidenced by:

Based on surveyor's review of the Cytology procedure manual, lack of workload records and an interview the pathologist/laboratory director/technical supervisor, the pathologist failed to record the number of hours spent examining the cytology slides from August 3, 2020 through July 14, 2021. FINDINGS: 1. The pathologist /laboratory director/technical supervisor confirmed on July 14, 2021 at approximately 1:30 PM, the surveyor's findings that the pathologist, as the primary reader, failed to record the total number of slides examined in a 24-hour period and the number of

hours spent examining slides from August 3, 2020 through July 14, 2021. 2. The pathologist stated, "that the number of slides screened and the hours screened, were not documented for this location."

D5647

CYTOLOGY
CFR(s): 493.1274(d)(4)

(d) Workload limits. The laboratory must establish and follow written policies and procedures that ensure the following: (d)(4) Records are available to document the workload limit for each individual.

This STANDARD is not met as evidenced by:
Based on surveyor's review of Cytology policies and procedures, laboratory records and confirmed in a interview with the pathologist/laboratory director/technical supervisor , at the time of this survey, the laboratory failed to establish written policies and procedures to ensure that records are maintained and available to document the workload for the individual who performs primary screening of non-gynecologic cytology slides

D6079

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1445(a)(b)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, record and report test results promptly, accurately and proficiently, and for assuring compliance with the applicable regulations. (a) The laboratory director, if qualified, may perform the duties of the technical supervisor, clinical consultant, general supervisor, and testing personnel, or delegate these responsibilities to personnel meeting the qualifications under 493.1447, 493.1453, 493.1459, and 493.1487 respectively. (b) If the laboratory director reapportions performance of his or her responsibilities, he or she remains responsible for ensuring that all duties are properly performed.

This STANDARD is not met as evidenced by:
Based on surveyor's review of laboratory policies and procedures, laboratory records and confirmed in an interview with the pathologist/laboratory director/technical supervisor, the laboratory director/technical supervisor failed to be responsible for the overall operation and administration of the laboratory, to include assuring compliance with the applicable regulations and ensuring that all the duties of the laboratory director were performed. Refer to D5217, D5291, D5437, D5601, D5631, D5633, D5637, D5645 and D5647

D6093

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1445(e)(5)

The laboratory director must ensure that the quality control programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.

This STANDARD is not met as evidenced by:

Based on review of laboratory policies and procedures, histopathology QC records and confirmed in an interview with the pathologist/laboratory director/technical supervisor, the laboratory director failed to ensure that the established histology and cytology quality control program was maintained to assure the quality of histology and cytology testing and identify failures. Refer to D5437 and D5601

D6094

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1445(e)(5)

The laboratory director must ensure that the quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.

This STANDARD is not met as evidenced by:
Based on surveyor's review of laboratory policies and procedures, QC and QA laboratory records and confirmed in an interview with the pathologist/laboratory director/ technical supervisor, the laboratory director/technical supervisor failed to ensure that quality assessment (QA) programs were established to assure the quality of laboratory services and identify failures in quality as they occur. Refer to D5291