

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 33D2072646 | (X3) Date Survey Completed 08/02/2022 |
| Name of Provider or Supplier Schweiger Dermatology, Pllc - Pathology Lab | Street Address, City, State 65 Broadway, Suite 1606, New York, NY | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D5801 | <p>TEST REPORT CFR(s): 493.1291(a)</p> <p>The laboratory must have an adequate manual or electronic system(s) in place to ensure test results and other patient-specific data are accurately and reliably sent from the point of data entry (whether interfaced or entered manually) to final report destination, in a timely manner. This includes the following: (a)(1) Results reported from calculated data. (a)(2) Results and patient-specific data electronically reported to network or interfaced systems. (a)(3) Manually transcribed or electronically transmitted results and patient-specific information reported directly or upon receipt from outside referral laboratories, satellite or point-of-care testing locations.</p> <p>This STANDARD is not met as evidenced by: Based on review of the response to the plan of correction from the compliant revisit on June 29, 2022, the review of the newly implemented Quality Assessment Policy, the laboratory failed to have a system in place to verify the accuracy of their LIS system transfer of data. FINDINGS: The laboratory failed to have a mechanism in place to periodically verify the accuracy of test results to ensure that the results are accurately sent from the point of data entry to the final report destination.</p> |