

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  33D2148508	<b>(X3) Date Survey Completed</b>  02/27/2019
<b>Name of Provider or Supplier</b>  Active Medical Care Pc	<b>Street Address, City, State</b>  275 North Middletown Road, Suite 1f, Pearl River, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5403</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(b)</p> <p>The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals (normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life threatening results, or panic, or alert values. (14) Description of the course of action to take if a test system becomes inoperable.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory procedure manual and an interview with the laboratory director and technical supervisor, the laboratory failed to have a complete transport procedure. Findings Include: The laboratory director and technical supervisor confirmed on February 27, 2019, at approximately 1:00 pm, the laboratory failed to have a tracking policy/procedure for specimens sent from the New Jersey offices to the New York office where patient specimen testing is performed.</p>

**D5805****TEST REPORT**

CFR(s): 493.1291(c)

The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.

This STANDARD is not met as evidenced by:

Based on a review of patient test reports and an interview with the laboratory director and technical supervisor, the laboratory did not include the date of the test report. Findings Include: The laboratory director and technical supervisor confirmed on February 27, 2018, at approximately 1:00 pm that the laboratory failed to have the test report date on four (4) of four (4) patient test reports.

**D6094****LABORATORY DIRECTOR RESPONSIBILITIES**

CFR(s): 493.1445(e)(5)

The laboratory director must ensure that the quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.

This STANDARD is not met as evidenced by:

Based on a review of procedures and an interview with the laboratory director and technical supervisor, the director failed to ensure that the laboratory's QA program was maintained as part of the laboratory's overall quality systems program. Refer to D5403 and D5805