

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D2181029	(X3) Date Survey Completed 10/12/2022
Name of Provider or Supplier Naushin Siddiqui Physician Pc	Street Address, City, State 4250 Hempstead Turnpike - Ste 14, Bethpage, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6019	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(iv)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(iv) Ensure that an approved corrective action plan is followed when any proficiency testing results are found to be unacceptable or unsatisfactory.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's review of the laboratory's API PT records and an interview with the lab director on 10/12/2022 about 1pm, the laboratory director failed to ensure that corrective action was performed and documented for the laboratory's unsatisfactory PT performance of 2022 first event immunology IGE 80%.</p>