

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D2231103	(X3) Date Survey Completed 12/12/2024
Name of Provider or Supplier James C Prezzano Md Pllc	Street Address, City, State 212 Highbridge St, Suite A, Fayetteville, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation, review of thermometer calibration records and Standard Operating Procedures (SOPs), as well as interview with the Testing Personnel (TP), the laboratory failed to perform and document thermometer calibration. FINDINGS: 1. The most recent calibration of the Thomas Scientific thermometer, Model: 9337T07, serial number (SN): 221228574, utilized for monitoring room temperature and humidity of the Mohs processing laboratory, expired February 8, 2024. There was no documentation of calibration performance since expiration. 2. The current, approved SOPs did not include instructions for performing such activity. 3. The TP confirmed the findings on December 12, 2024, at approximately 4:00 P.M.</p>