

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 33D2272658	(X3) Date Survey Completed 07/14/2023
Name of Provider or Supplier Long Island Dermatology	Street Address, City, State 1615 Northern Blvd Suite 400, Manhasset, NY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation of the flammable materials safety cabinet contents, the laboratory failed to remove expired reagents from inventory. Findings: 1. One bottle of Scott's Tap Water Substitute; Lot #: 2215119; Exp. June 10, 2023, was in inventory. 2. Confirmed finding by interview with the office manager on July 14, 2023, at approximately 11:00 A.M.</p>