

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  34D0237483	<b>(X3) Date Survey Completed</b>  03/19/2026
<b>Name of Provider or Supplier</b>  Atrium Health Wfb Family Medicine Laboratory	<b>Street Address, City, State</b>  1920 West First Street, Winston Salem, NC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the TC (technical consultant) 3/19/26, the laboratory failed to discard supplies that exceeded their expiration date. During a tour of the laboratory at approximately 12:40 p.m., the surveyor observed 9 Vacuette yellow top tubes (lot #A2409454, expiration date 3/15/26) on the counter, available for use. During interview at approximately 12:45 p.m., the TC confirmed the tubes were expired. She stated the laboratory had used some of the expired tubes to send patient specimens to an outside laboratory for testing.</p>